



FFP2 PROTECTIVE RESPIRATORS

Protective respirators with 4layer filtration system.
DM-002(Without air valve), DM-002V (With air valve).
Available in 2 or 10pcs/pack.

SPECIFICATION:

Standard: CE0161, EN149:2001+A1:2009, COVID-19 PPE-R/02.075-V2

Style: Strap behind ears/Behind the wearer's head (includes head strip holder)

Filtration efficiency: Above 95%K (Equivalent to US N95 and EU FFP2 standard)

Marking: PPE marking, FFP2NR, CE 0161, EN149:2001+A1:2009, COVID-19 PPE-R/02.075-V2

Designation: NR-Non reusable (single shift use only)





EU Declaration of Conformity

REGULATION (EU) 2016/425 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 9 March 2016 on personal protective equipment and repealing Council Directive 89/686/ EEC

This is hereby declared that following designated product complied with the essential health and safety requirements of above Council Directive (s) on the approximation of the laws of the Member State relating to it

MANUFACTURER:	TECHDENTIKA Tilzes st. 86, Siauliai, Lithuania
PRODUCT	FOLD FLAT MASK
PRODUCT NAME:	DMC MASK
CLASSIFICATION:	FFP2
MODEL:	DM-002 (fold flat mask)
CERTIFICATE:	EN149:2001+A1:2009 CERTIFICATE No. 21/3782/00/0161
CONFORMITY ASSESSMENT PROCEDURE ACORDING TO (EU) 2016/425, PPE-R02.275 VERSION2:	EU Type Examination Certificate Number: Test report 2020EC4785
NOTIFIED BODY INFORMATION:	INVESTIGACION DE LA INDUSTRIA TEXTIL AITEX, Plaza Emilio Sala 1, Spain CE 0161
EVALUATION OF THE CONFORMITY:	2021EC0051UE

DM-002 is in conformity with the provisions of Regulation (EU) 2016/425, including fulfilment of the applicable essential health and safety requirements set out in Annex II, and with the National Standard transposing the harmonized European Standard Number EN 149:2001+A1:2009. DM-002 is subject to the conformity assessment procedure Module C2 under surveillance of the notified body ASSOCIATION OF RESEARCH OF THE TEXTILE INDUSTRY, notified body No 0161.

Signed on behalf of **TECHDENTIKA**

VILNIUS, LITHUANIA
Date: 2021.01.27
General manager VYTAUTAS ADOMAITIS



This declaration is the responsibility of the Manufacturer

Document valid until 2022 01 27.





CERTIFICATE No. 21/3782/00/0161

PPE TYPE FILTERING HALF MASK FOR COVID-19 PROTECTION ONLY
REF: DM-002



AITEX, Notified Body No. 0161 for the application of Regulation (EU) 2016/425 of the European Parliament and of the Council of 9th March 2016, in which the essential health and safety requirements that Personal Protective Equipment (PPE) must comply with.

CERTIFIES The Company:

UAB TECHDENTIKA
SVITRIGAILOS 11B
LT-03228 VILNIUS

As a manufacturer



CERTIFICATE No. 21/3782/00/0161

Has obtained EU TYPE EXAMINATION in compliance with what is set out in Annex V (module B) in Regulation (EU) 2016/425 and in agreement with the applicable test procedures and technical specifications

Respiratory protection against COVID-19 only, according to the following standard/s:

- PPE-R / 02.075 version 2. Filter half mask to protect against COVID-19 only on the basis of Commission Recommendation (EU) 2020/403 of March 13, 2020 on conformity assessment and procedures for Market surveillance in the context of the threat posed by COVID-19.

Having achieved the performance requirements specified in conformity assessment report No. **2021EC0051UE** and the PPE's Technical Documentation.

Description of the PPE:

Particle filtering half mask without exhalation valve covering nose, mouth and chin, white colour.

The particle filtering half mask has ear bands and nose clip.

The materials that form the PPE, are described in the conformity assessment report nº 2021EC0051UE.

It shall be the manufacturer's responsibility to provide specific information of this certificate and the tested levels of protection.

The CAT. III PPE shall only be used in conjunction with one of the conformity assessment procedures according to module C2 or module D described in article 19 letter c) of the Regulation (EU) 2016/425.

Digitally Signed by: Silvia Devesa

Date: 27/01/2021 11:03:02

Location: Alcoy

Silvia Devesa Valencia
Laboratory Subdirector and Innovation

Date of issue of the Certificate: 26th of January 2021

Date of expiry: 26th of January 2022